

Soil quality —

Determination of carbonate content — Volumetric method

ICS 13.080.10

National foreword

This British Standard is the UK implementation of EN ISO 10693:2014. It is identical to ISO 10693:1995. It supersedes BS 7755-3.10:1995 (dual numbered as ISO 10693:1995), which is withdrawn.

The UK participation in its preparation was entrusted to Technical Committee EH/4, Soil quality.

A list of organizations represented on this committee can be obtained on request to its secretary.

Additional information. References to labelled parts of the apparatus refer exclusively to Figure 1.

In paragraph 1 of 7.2 it is recommended that the variation in temperature is no more than $\pm 1\text{ }^{\circ}\text{C}$ and the variation in pressure is no more than $\pm 0.5\text{ kPa}$.

In paragraph 2 of 7.2 the phrase “two standards of respectively 0,200 g and 0,400 g” means that one sample of each should be prepared.

Textual errors. In note 2 in clause 3, “Annex A ” should read “Annex B”. In paragraph 4 of 7.2 “tube H” should read “tube E”.

This publication does not purport to include all the necessary provisions of a contract. Users are responsible for its correct application.

Compliance with a British Standard cannot confer immunity from legal obligations.

This British Standard, having been prepared under the direction of the Health and Environment Sector Board, was published under the authority of the Standards Board and comes into effect on 15 October 1995

© The British Standards Institution 2014. Published by BSI Standards Limited 2014

ISBN 978 0 580 83738 8

Amendments/corrigenda issued since publication

Date	Comments
30 June 2014	This corrigendum renumbers BS 7755-3.10:1995 (dual numbered as ISO 10693:1995) as BS EN ISO 10693:2014