

CORRIGENDUM 1

201.3.201

Replace the existing term " $B_{1\text{RMS}}$ " by " $B_{1+}\text{RMS}$ ".

Replace, in the definition:

*" B_1 , the radio frequency magnetic induction"
by*

*" B_{1+} , the MR relevant radiofrequency
magnetic induction"*

*Replace, in the formula, " $B_{1\text{RMS}}$ " by " $B_{1+}\text{RMS}$ "
and " B_1 " by " B_{1+} ".*

201.3.201

*Remplacer le terme existant " $B_{1\text{RMS}}$ " par
" $B_{1+}\text{RMS}$ ".*

Remplacer, dans la définition:

*" B_1 , l'induction magnétique à radiofréquence"
par*

*" B_{1+} , l'induction magnétique à radiofréquence
RM pertinente".*

*Remplacer, dans la formule, " $B_{1\text{RMS}}$ " par
" $B_{1+}\text{RMS}$ " et " B_1 " par " B_{1+} ".*

Table 201.101 – List of symbols

*In the 2nd row, 1st column replace " B_1 " by
" B_{1+} ".*

*In the 3rd row, 1st column replace " $B_{1\text{RMS}}$ " by
" $B_{1+}\text{RMS}$ ".*

*In the 3rd row, 3rd column replace " B_1 " by
" B_{1+} ".*

Tableau 201.101 – Liste des symboles

*Dans la 2^{ème} ligne, 1^{ère} colonne remplacer
" B_1 " par " B_{1+} ".*

*Dans la 3^{ème} ligne, 1^{ère} colonne remplacer
" $B_{1\text{RMS}}$ " par " $B_{1+}\text{RMS}$ ".*

*Dans la 3^{ème} ligne, 3^{ème} colonne remplacer
" B_1 " par " B_{1+} ".*

201.7.9.2.101 Instructions for use for MR EQUIPMENT

j) Exposure of the PATIENT to radio
frequency magnetic fields

*In the final dashed item replace " $B_{1\text{RMS}}$ " by
" $B_{1+}\text{RMS}$ "*

201.7.9.2.101 Instructions d'utilisation des APPAREILS A RM

j) Exposition du PATIENT aux champs
magnétiques à radiofréquence

*Dans le dernier tiret, remplacer " $B_{1\text{RMS}}$ " par
" $B_{1+}\text{RMS}$ ".*