



INTERNATIONAL STANDARD 6721-3:1994
TECHNICAL CORRIGENDUM 1

Published 1995-03-15

INTERNATIONAL ORGANIZATION FOR STANDARDIZATION • МЕЖДУНАРОДНАЯ ОРГАНИЗАЦИЯ ПО СТАНДАРТИЗАЦИИ • ORGANISATION INTERNATIONALE DE NORMALISATION

Plastics — Determination of dynamic mechanical properties —

Part 3: Flexural vibration — Resonance-curve method

TECHNICAL CORRIGENDUM 1

Plastiques — Détermination des propriétés mécaniques dynamiques —

Partie 3: Vibration en flexion — Méthode en résonance

RECTIFICATIF TECHNIQUE 1

Technical corrigendum 1 to International Standard ISO 6721-3:1994 was prepared by Technical Committee ISO/TC 61, *Plastics*, Subcommittee SC 2, *Mechanical properties*.

Page 5

Subclause 9.6

Correct the first sentence in the second paragraph to read:

“Measure the amplitude to $\pm 0,5$ %, the resonance frequency to at least $\pm 0,1$ % and the width of the resonance peaks to ± 1 % of the value of the peak width (see 11.2).”

ICS 83.080.00

Ref. No. ISO 6721-3:1994/Cor.1:1995(E)

Descriptors: plastics, rigid plastics, tests, vibration tests, determination, mechanical properties, dynamic properties, modulus of elasticity, test equipment.

© ISO 1995

Printed in Switzerland

INTERNATIONAL STANDARD

ISO
6721-3

First edition
1994-11-01

Plastics — Determination of dynamic mechanical properties —

Part 3:

Flexural vibration — Resonance-curve method

*Plastiques — Détermination des propriétés mécaniques dynamiques —
Partie 3: Vibration en flexion — Méthode en résonance*



Reference number
ISO 6721-3:1994(E)